



Glassware List

Category: - Volumetric Glassware

S. No.	Item Name	Capacity ml	Quantity/ Pack Size		Picture
1	Flask Boiling, Round Bottom, Plain	25	2		
2		50	2		
3		100	2		
4		150	2		
5		250	2		
6		500	2		
7		1000	1		

S. No.	Item Name	Capacity ml	Quantity/ Pack Size		Picture
1	Flask Boiling, Flat Bottom, Plain	25	2		
2		50	2		
3		100	2		
4		150	2		
5		300	2		
6		250	2		
7		500	1		
8		1000	1		


S. No.	Item Name	Capacity ml	Quantity/ Pack Size		Picture
1	Kjeldahl Flask with I.C. Joint	800	1		
2		800	1		

S. No.	Item Name	Capacity ml	Graduation	Tolerance	Quantity/ Pack Size		Picture
1	Burettes with Straight Bore Glass Key with Screw Thread, O-ring Stopcock, Accuracy As Per Class-B, of I.S :1997-2008,ISO 385:2005	10	0.05	0.05	1		
2		25	0.1	0.1	1		
3		50	0.1	0.1	1		
4		100	0.2	0.2	1		

S. No.	Item Name	Capacity ml	Graduation	Tolerance	Quantity/ Pack Size		Picture
1	Burette with Straight Bore, Rotaflow Screw Thread Stopcock, PTFE Keys, Accuracy As Per Class-A of IS-1997:2008, ISO/DIN 385:2005, with Certificate.	10	0.05	0.02	1		
2		25	0.1	0.05	1		
3		50	0.1	0.05	1		
4		100	0.2	0.1	1		
S. No.	Item Name	Capacity ml	Graduation	Tolerance	Quantity/ Pack Size		Picture
1	Burettes Straight Bore Rotaflow Stopcock Class-B, Standard, Screw Thread Stopcocks, PTFE Keys, Accuracy as Per Class-B, of IS-1997:2008, ISO/DIN 385:2005	10	0.05	0.05	1		
2		25	0.1	0.1	1		
3		50	0.1	0.1	1		
4		100	0.2	0.2	1		
S. No.	Item Name	Capacity ml	Graduation	Tolerance	Quantity/ Pack Size		Picture
1	Burettes with Straight Bore Glass Key with Screw Thread, O-ring Stopcock, Accuracy as Per Class-A of I.S. :1997-2008 with Work Certificate, Complies with IS 1997, ISO/DIN 385:2005	10	0.05	0.02	1		
2		25	0.1	0.05	1		
3		50	0.1	0.05	1		
4		100	0.2	0.1	1		
S. No.	Item Name	Capacity ml	Graduation	Tolerance	Quantity/ Pack Size		Picture
1	Burettes Straight Bore, PTFE Key Stopcock, Accuracy As Per Class-B, of I.S. 1997:2008/ISO 385:2005	10	0.05	0.05	1		
2		25	0.1	0.1	1		
3		50	0.1	0.1	1		
4		100	0.2	0.2	1		

S. No.	Item Name	Capacity ml	Graduation	Tolerance	Quantity/ Pack Size		Picture
1	Burettes Straight Bore, PTFE Key Stopcock, Accuracy As Per Class-A, of I.S. 1997:2008/ISO 385:2005, with Certificate	10	0.05	0.02	1		
2		25	0.1	0.05	1		
3		50	0.1	0.05	1		
4		100	0.2	0.1	1		
S. No.	Item Name	Capacity ml	Tolerance	Bottle Capacity(ml)	Quantity/ Pack Size		Picture
1	Automatic Burette Zero with PTFE Needle Valve Stopcock, Complies With Din ISO 385 Standard Class-B	10	0.05	500	1		
2		25	0.1	2000	1		
3		50	0.1	2000	1		
4		100	0.2	2000	1		
S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Nessler Cylinder For Colour Comparison with Class-A, Graduated As Per IS : 4161:1967, with Certificate	50	25 to 50	0.4	2		
2		100	50 to 100	0.8	2		
S. No.	Item Name	Capacity ml	Tolerance	Bottle Capacity (ml)	Quantity/ Pack Size		Picture
1	Burette, Automatic Zero with PTFE Needle Valve Stopcock. Complies with Din ISO 385 Standard Class-A, comes with Individual Works Certificate, with Detachable Stopcock Stopper	10	0.03	500	1		
2		25	0.05	2000	1		
3		50	0.05	2000	1		
4		100	0.1	2000	1		

S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Measuring Cylinder, with Spout, Round Base, Accuracy as Per Class-B of IS 878 :2008 ISO 4788 : 2005	5	0.1	0.1	2		
2		10	0.2	0.2	2		
3		25	0.5	0.5	2		
4		50	1.0	1.0	2		
5		100	1.0	1.0	2		
6		250	2.0	2.0	2		
7		500	5.0	5.0	2		
8		1000	10.0	10.0	1		
9		2000	20.0	20.0	1		

S. No.	Item Name	Capacity ml	Graduation Interval	Quantity/ Pack Size		Picture
1	Nessler Cylinder For Colour Comparison with Class-B, Graduated As Per Is : 4161:1967	50	25 to 50	2.0		
2		100	50 to 100	2.0		

S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Measuring Cylinder, with Spout, Round Base, Accuracy as Per Class-A, of IS 878 : 2008 ISO 4788 : 2005, With Certificate.	5	0.1	0.05	2		
2		10	0.2	0.1	2		
3		25	0.5	0.3	2		
4		50	1.0	0.5	2		
5		100	1.0	0.5	2		
6		250	2.0	1.0	2		
7		500	5.0	2.5	2		
8		1000	10.0	5.0	1		
9		2000	20.0	10.0	1		

S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Measuring Cylinder, Round Base, Graduated,with Interchangeable Stopper, Class-B, Accuracy AS IS 878 : 2008, ISO 4788 : 2005	5	0.1	0.1	2		
2		10	0.2	0.2	2		
3		25	0.5	0.5	2		
4		50	1.0	1.0	2		
5		100	1.0	1.0	2		
6		250	2.0	2.0	2		
7		500	5.0	5.0	2		
8		1000	10.0	10.0	1		
9		2000	20.0	20.0	1		

S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Measuring Cylinder, Round Base, Graduated, with Interchangeable Stopper, Class-A, Accuracy As IS 878 : 2008, ISO 4788 : 2005 with Certificate	5	0.1	0.05	2		
2		10	0.2	0.1	2		
3		25	0.5	0.3	2		
4		50	1.0	0.5	2		
5		100	1.0	0.5	2		
6		250	2.0	1.0	2		
7		500	5.0	2.5	2		
8		1000	10.0	5.0	1		
9		2000	20.0	10.0	1		

S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Cylinders, Graduated, with Spout, Hexagonal Base, Accuracy as Per Class-A, of IS 878 : 2008, ISO/ DIN 4788:2005, with Certificate	5	0.1	0.05	2		
2		10	0.2	0.1	2		
3		25	0.5	0.25	2		
4		50	1.0	0.5	2		
5		100	1.0	0.5	2		
6		250	2.0	1.0	2		
7		500	5.0	2.5	2		
8		1000	10.0	5.0	1		
9		2000	20.0	10.0	1		


S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Cylinders, Graduated, with Spout, Hexagonal Base, Accuracy as Per Class-B, of IS 878 : 2008, ISO/DIN 4788:2005	5	0.1	0.05	2		
2		10	0.2	0.1	2		
3		25	0.5	0.25	2		
4		50	1.0	0.5	2		
5		100	1.0	0.5	2		
6		250	2.0	1.0	2		
7		500	5.0	2.5	2		
8		1000	10.0	5.0	1		
9		2000	20.0	10.0	1		



S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Cylinders, Graduated, with Spout, Hexagonal Base, Accuracy as Per Class-A, USP / ASTM E-1272, with Certificate	5	0.1	0.05	2		
2		10	0.2	0.1	2		
3		25	0.5	0.25	2		
4		50	1.0	0.5	2		
5		100	1.0	0.5	2		
6		250	2.0	1.0	2		
7		500	5.0	2.5	2		
8		1000	10.0	5.0	1		
9		2000	20.0	10.0	1		
S. No.	Item Name	Capacity ml	Tolerance	Stopper Size	Quantity/ Pack Size		Picture
1	Flasks, Volumetric, with Interchangeable Glass Stopper, Accuracy as Per Class- B, of IS 915:2012,ISO/DIN 1042	1	0.05	Size 10/19	2		
2		2	0.05	Size 10/19	2		
3		5	0.05	Size 10/19	2		
4		10	0.05	Size 10/19	2		
5		20	0.08	Size 10/19	2		
6		25	0.08	Size 10/19	2		
7		50	0.12	Size 12/21	2		
8		100	0.2	Size 14/23	2		
9		150	0.2	Size 14/23	2		
10		200	0.3	Size 14/23	2		
11		250	0.3	Size 14/23	2		
12		500	0.5	Size 19/26	2		
13		1000	0.8	Size 19/26	2		
14		2000	1.2	Size 24/29	2		
S. No.	Item Name	Capacity ml	Tolerance	Stopper Size	Quantity/ Pack Size		Picture
1	Flask, Volumetric, with Interchangeable Glass Stopper, Accuracy As Per Class-A, of IS 915 : 2012, ISO/DIN 1042, With Certificate	1	0.025	Size 10/19	2		
2		2	0.025	Size 10/19	2		
3		5	0.025	Size 10/19	2		
4		10	0.025	Size 10/19	2		
5		20	0.04	Size 10/19	2		
6		25	0.04	Size 10/19	2		
7		50	0.06	Size 12/21	2		
8		100	0.1	Size 14/23	2		
9		150	0.1	Size 14/23	2		
10		200	0.15	Size 14/23	2		
11		250	0.15	Size 14/23	2		
12		500	0.25	Size 19/26	2		
13		1000	0.4	Size 19/26	2		
14		2000	0.6	Size 24/29	2		

S. No.	Item Name	Capacity ml	Tolerance	Stopper Size	Quantity/ Pack Size		Picture
1	Flask, Volumetric, with Interchangeable Glass Stopper, Accuracy As Per Class-A, USP & ASTM E288, With Certificate	5	0.02	Size 10/19	2		
2		10	0.02	Size 10/19	2		
3		25	0.03	Size 10/19	2		
4		50	0.05	Size 10/19	2		
5		100	0.08	Size 14/23	2		
6		200	0.1	Size 14/23	2		
7		250	0.12	Size 14/23	2		
8		500	0.2	Size 19/26	2		
9		1000	0.3	Size 19/26	2		
S. No.	Item Name	Capacity ml	Tolerance	Stopper Size	Quantity/ Pack Size		Picture
1	Flask, Volumetric, Amber Colour, with Interchangeable Glass Stopper, Accuracy As Per Class-B, of IS 915:2012, ISO / DIN 1042	1	0.05	Size 10/19	2		
2		2	0.05	Size 10/19	2		
3		5	0.05	Size 10/19	2		
4		10	0.05	Size 10/19	2		
5		20	0.08	Size 10/19	2		
6		25	0.08	Size 10/19	2		
7		50	0.12	Size 12/21	2		
8		100	0.2	Size 14/23	2		
9		200	0.3	Size 14/23	2		
10		250	0.3	Size 14/23	2		
11		500	0.5	Size 19/26	2		
12		1000	0.8	Size 19/26	2		
13		2000	1.2	Size 24/29	2		
S. No.	Item Name	Capacity ml	Tolerance	Stopper Size	Quantity/ Pack Size		Picture
1	Flask, Volumetric, Amber Colour, with Interchangeable, Glass Stopper, Accuracy As Per Class-A, of IS 915:2012, ISO / DIN 1042	1	0.025	Size 10/19	2		
2		2	0.025	Size 10/19	2		
3		5	0.025	Size 10/19	2		
4		10	0.025	Size 10/19	2		
5		20	0.04	Size 10/19	2		
6		25	0.04	Size 10/19	2		
7		50	0.06	Size 12/21	2		
8		100	0.1	Size 14/23	2		
9		150	0.1	Size 14/23	2		
10		200	0.15	Size 14/23	2		
11		250	0.15	Size 14/23	2		
12		500	0.25	Size 19/26	2		
13		1000	0.4	Size 19/26	2		
14		2000	0.6	Size 24/29	2		

S. No.	Item Name	Capacity ml	Tolerance	Quantity/ Pack Size	Price/ Piece Rs.	Picture
1	Pipettes Transfer Volumetric, with One Mark, Accuracy As Per Class-B, of IS :1117 :1975 / ISO DIN 648	1	0.015	10		
2		2	0.02	10		
3		5	0.03	10		
4		10	0.04	10		
5		20	0.06	10		
6		25	0.06	10		
7		50	0.1	10		
8		100	1.16	10		

S. No.	Item Name	Capacity ml	Tolerance	Quantity/ Pack Size		Picture
1	Pipettes Transfer Volumetric, with One Mark, Accuracy As Per Class-A, of IS :1117 :1975 / ISO DIN 648, With Certificate	1	0.006	10		
2		2	0.01	10		
3		3	0.01	10		
4		4	0.01	10		
5		5	0.015	10		
6		6	0.015	10		
7		7	0.015	10		
8		8	0.02	10		
9		9	0.02	10		
10		10	0.02	10		
11		15	0.02	10		
12		20	0.03	10		
13		25	0.03	10		
14		50	0.05	10		
15		100	0.08	10		

S. No.	Item Name	Capacity ml	Tolerance	Quantity/ Pack Size		Picture
1	Pipette, Transfer Volumetric, Accuracy As Per Clas-A, USP/ASTM Standard, E969-02, with Certificate	1	0.006	10		
2		2	0.006	10		
3		3	0.01	10		
4		4	0.01	10		
5		5	0.01	10		
6		6	0.01	10		
7		7	0.01	10		
8		8	0.02	10		
9		9	0.02	10		
10		10	0.02	10		
11		15	0.03	10		
12		20	0.03	10		
13		25	0.03	10		
14		30	0.03	10		
15		40	0.05	10		
16		50	0.05	10		
17		100	0.08	10		


S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Pipettes, Measuring, Graduated, Serological, Colour Code, Accuracy As Per Class- B, of IS / ISO 835 : 2007	0.1	0.01	0.01	10		
2		0.2	0.01	0.01	10		
3		0.5	0.02	0.02	10		
4		1	0.01	0.01	10		
5		2	0.02	0.02	10		
6		5	0.05	0.05	10		
7		10	0.1	0.1	10		
8		25	0.2	0.2	10		
S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Pipettes, Measuring, Graduated, Serological, Colour Code, Accuracy As Per Class-A, of IS / ISO 835 : 2007, With Certificate	0.1	0.01	0.006	10		
2		0.2	0.01	0.006	10		
3		0.5	0.02	0.006	10		
4		1	0.01	0.007	10		
5		2	0.02	0.01	10		
6		5	0.05	0.03	10		
7		10	0.1	0.05	10		
8		25	0.02	0.1	10		
S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Pipettes, Measuring (Mohr Type), Colour Code, Accuracy As Per Class-B, of IS/ISO/DIN 835 :2007	0.1	0.01	0.01	10		
2		0.2	0.01	0.01	10		
3		0.5	0.02	0.01	10		
4		1	0.01	0.01	10		
5		2	0.02	0.02	10		
6		5	0.5	0.1	10		
7		10	0.1	0.1	10		
8		25	0.2	0.2	10		

S. No.	Item Name	Capacity ml	Graduation Interval	Tolerance	Quantity/ Pack Size		Picture
1	Pipettes, Mesuring (Mohr Type), Colour Code, Accuracy As Per Class-A, of IS/ISO/DIN 835 : 2007 With Certificate	0.1	0.01	0.006	10		
2		0.2	0.01	0.006	10		
3		0.5	0.02	0.006	10		
4		1	0.01	0.007	10		
5		2	0.02	0.01	10		
6		5	0.05	0.03	10		
7		10	0.1	0.05	10		
8		25	0.2	0.1	10		

S. No.	Item Name	Capacity ml	Tolerance	Quantity/ Pack Size		Picture
1	Crow Receiver, Class-B	50	0.2 & 0.4	2		
2		100	0.4 & 0.8	2		

S. No.	Item Name	Quantity/ Pack Size			Picture
1	Pipettes for Milk Testing 10.75 ML, As Per I.S. Specification No.1223, Part II : 1972, For Class-A, with Certificate	10			

S. No.	Item Name	Capacity ml	Tolerance	Quantity/ Pack Size		Picture
1	Crow Receiver, Class-A, With Certificate	50	0.1 & 0.2	2		
2		100	0.2 & 0.4	2		

S. No.	Item Name	Capacity ml	Tolerance	Quantity/ Pack Size		Picture
1	Crow Receiver, Round Base, with Interchangeable Stopper, Class-B	50	0.2 & 0.4	2		
2		100	0.4 & 0.8	2		

S. No.	Item Name	Capacity ml	Tolerance	Quantity/ Pack Size		Picture
1	Crow Receiver, Round Base, with Interchangeable Stopper, Class-A, with Certificate	50	0.1 & 0.2	2		
2		100	0.2 & 0.4	2		

S. No.	Item Name	Extractor Cap (ml)	Extractor Socket	Extractor Cone	Flask Cap ml		Picture
1	Soxhlet Extraction, Consists of Flask Extractor and Condenser	60	34/35	24/29	150		
2		100	34/35	24/29	250		
3		100	40/38	24/29	250		
4		200	50/42	24/29	500		

S. No.	Item Name	Extractor Cap (ml)	Extractor Socket	Extractor Cone		Picture
1	Soxhlet Extractor, Only For Cat No.149	60	34/35	24/29		
2		100	34/35	24/29		
3		100	40/38	24/29		
4		200	50/42	24/29		

S. No.	Item Name	Capacity ml	Quantity/ Pack Size		Picture		
1	Specific Gravity Bottles, with Capillary Teflon Stopper	5	12				
2		10	12				
3		25	12				
4		50	12				
5		100	6				
S. No.	Item Name	Capacity ml	Tolerance		Picture		
1	Specific Gravity Bottles, with Capillary Bore Interchangeable Teflon Stopper, Class-A, with Works Certificate	5	0.1				
2		10	0.3				
3		25	0.8				
4		50	1.0				
5		100	2.0				
S. No.	Item Name	TALL FORM / SQUATE FORM	O.D. X HT (mm)	Capacity		Picture	
1	Weighing Bottles, With Interchangeable Stopper	TALL FORM	20X40	5			
2		TALL FORM	25X60	15			
3		TALL FORM	30X60	25			
4		TALL FORM	40X90	60			
5		SQUATE FORM	40X30	20			
6		SQUATE FORM	50X25	15			
7		SQUATE FORM	50X35	25			
8		SQUATE FORM	50X50	50			
9		SQUATE FORM	60X40	40			

S. No.	Item Name	Socket	Cone	Effective Length(mm)		Picture
1	Condensers, Liebig, Drip Tip Interchangeable Inner & Outer Joint	19/26	19/26	200		
2		24/29	24/29	300		
3		24/29	24/29	400		
4		24/29	24/29	500		
5		24/29	24/29	600		

S. No.	Item Name	Socket	Cone	Effective Length(mm)		Picture
1	Allihn Condensers (Bulb Condenser)	19/26	19/26	200		
2		24/29	24/29	300		
3		24/29	24/29	400		
4		24/29	24/29	500		

S. No.	Item Name	Socket	Cone	Effective Length(mm)		Picture
1	Coil Condensers, Graham, Jacket Distillate Type	19/26	19/26	200		
2		24/29	24/29	300		
3		24/29	24/29	400		
4		24/29	24/29	500		
5		24/29	24/29	600		

S. No.	Item Name	Capacity ml	Socket Size	Quantity/ Pack Size		Picture
1	Flask Iodine, With Funnel Shaped Cup & Stopper	100	24/29	2		
2		250	24/29	2		
3		500	24/29	2		

S. No.	Item Name	Capacity ml	Quantity/ Pack Size		Picture
1	B.O.D. Bottles,with Interchangeable Stopper and Individual Number	60	2		
2		125	2		
3		300	2		